

CERTIFICATE OF ANALYSIS

Product Name	Growth hormone-releasing hormone-like peptide 1(Cat.#304747)			
Lot No.	GTA34368-9-0712			
Sequence	HSDGIFTKDYRKYLGQLRAQKFLQWLM			
Dissolution condition	100%H2O			
Length	27AA			
Modification	N/A			
Molecular Weight (MW)	3343.86			
Storage	-20°C			
Test Items	Specifications	Results		
Test Items MW by MS	Specifications 3344.10	Results Conforms		
	-			
MW by MS	3344.10	Conforms		
MW by MS Purity by HPLC	3344.10 >95%	Conforms 95.036%		
MW by MS Purity by HPLC Peptide Content	3344.10 >95% N/A	Conforms 95.036% N/A		
MW by MS Purity by HPLC Peptide Content Moisture content	3344.10 >95% N/A N/A	Conforms 95.036% N/A N/A		

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:

Quality Assurance Department

Aug/01/2022






Sample Information

Order ID Name	:GTA34368-9 :Growth hormone-relev	sing hormone-like penti	1 <u></u> 1		
Inallic	:Growth hormone-releasing hormone-like peptide 1 (from Chelyosoma productum, sea squirt)				
Sequence	:HSDGIFTKDYRKYLGQLRAQKFLQWLM				
Lot. No	:GTA34368-9-0712				
Pump A	:0.1% Trifluoroacetic in 100% water				
Pump B	:0.1% Trifluoroacetic in 100% acetonitrile				
Total Flow	:1ml/min				
Wavelength	:214nm				
Analytial column type	:YMC-Triart C18(4.6*250mm*5um)				
Dissolution method	:100%H2O				
Inj. Volume	: 35ul				
Time	Module	Action	Value		
0.01	Pumps	B.Conc	25		
30.00	Pumps	B.Conc	55		
33.00	Pumps	B.Conc	100		
38.00	Pumps	B.Conc	100		
40.00	Pumps	B.Conc	25		
50.00	Controller	Stop			

Chromatogram



Peak Table

Detector A 21	<u>4nm</u>			
Peak#	Ret. Time	Area	Height	Area%
1	16.587	92906	10572	2.218
2	16.889	3980753	568170	95.036
3	17.079	108711	14443	2.595
4	20.012	3584	463	0.086
5	20.657	2721	368	0.065
Total		4188674	594017	100.000



